

Measuring $\delta^{13}\text{C}$ and $\delta^{18}\text{O}$ of CO_2 at **Harvard Forest** using a Quantum Cascade Laser Spectrometer

*R. A. Wehr¹, J. W. Munger², D. D. Nelson³, J. B. McManus³,
M. S. Zahniser³, S. R. Saleska¹*



1. Ecology and Evolutionary Biology, University of Arizona, Tucson, AZ, United States
2. School of Engineering and Applied Sciences, Harvard University, Cambridge, MA, United States
3. Aerodyne Research Inc., Billerica, MA, United States



Calibration

MATERIALS: Stainless steel fittings, Synflex 1300 tubing, steel tanks → aluminum


FREQUENCY: Every 40 minutes

TYPE: 2-point linear [by isotopologue](#) (3-point linear for propagating tanks)

STANDARDS: Ordinary tanks, $[\text{CO}_2]$ measured at Harvard, deltas by SIRFER (by flask)



Data Output



SAMPLING FREQ:

4Hz (true bandpass 0.7 Hz)

AVERAGING:

Effectively 30 min for EC

(we take advantage of both

high freq and averaging)

90 s for standards (Allan min at 60 s)

SAVING SPECTRA:


So far, only every 100th saved, for diagnostics, due to volume of data and slow transfer times

DATA TRANSFER:



Mixing ratios, housekeeping by internet

Spectra will need portable drive

Data (Post-)Processing (in Igor)



Data Quality



SPECTROSCOPIC PROBLEM: Dependence of mixing ratios on room temperature
(buried in signal variability except on tanks)




Data Quality

NON-SPECTROSCOPIC PROBLEMS:


1. Clock drift and jumps between sonic anemometer and spectrometer
2. (Initially) pressure spikes, regulator instability
3. Purge air (for the optics), shack a/c breakdowns
4. Sonic anemometer breakdowns, bad data
5. Gas tank supply delays

INTRIGUE:

Strange isotopic cospectra and ogives sometimes



But... Results!



But... Results!

